

The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.

Effective February 5, 2024, Retractable Technologies, Inc. is recalling **lot K220402 EasyPoint Needle** due to the possible detachment of the needle cannula from the needle holder. The use of an EasyPoint Needle from lot K220402 may result in a serious injury if the needle cannula detaches from the needle holder in the patient.

PRODUCT NAME: EasyPoint Needle

LOT NUMBER: K220402

EXPIRATION DATES: 3/28/2027

DISTRIBUTION DATES: 7/20/2022 - 9/20/2023