

The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.

Description	Lot #	Exp Date	NDC	UPC
ONGUARD SPK PRT ADPT 412143 80	UBE517	06/30/2024	08021412143	72981191214
	UBE508	04/30/2022		
	UBE509	02/29/2024		
	UBE511	02/29/2024		
	UBE512	02/29/2024		
	UBE513	05/31/2024		
	UBE514	05/31/2024		
	UBE507	03/31/2022		
	UBE516	05/31/2024		
	UBE536	01/31/2025		
	UBE518	06/30/2024		
	UBE526	09/30/2024		
	UBE530	10/31/2024		
	UBE531	10/31/2024		
	UBE533	01/31/2025		
	UBE534	11/30/2024		
	UBE535	01/31/2025		
UBE515	05/31/2024			

Simplivia is recalling the above item/lot(s) due to 17 complaints associated with disconnection of the OnGuard®2 Bag Adaptor SP from infusion bags. This recall is to the retail level. Affected product started shipping September 10, 2020.