

*The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.*

B. Braun Medical Inc. (BBMI) is issuing a product recall for 0.9% Sodium Chloride Injection USP in Excel Plus IV Container, 1000 ml due to potential for bags to leak compromising the sterility of the product. The depth of this recall is to the user/consumer level.

The lots were inadvertently distributed to the market prior to completion of the required acceptance activities. In particular, planned acceptance activities for container integrity testing was not yet complete.

Product Catalog Number:	NDC Number	Product Description:	Lot Number	Distribution Range	Expiration Date	Region Distributed
Q8000	0264-5802-00	0.9% Sodium Chloride Injection USP in Excel Plus IV Container, 1000 ml	0061858305	01 Feb 2023 - 11 Apr 2023	31 Mar 2025	US
Q8000	0264-5802-00	0.9% Sodium Chloride Injection USP in Excel Plus IV Container, 1000 ml	0061858306	11 Apr 2023 - 12 Apr 2023	31 Mar 2025	US