

The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.

REF Number	ICC	Batch	Description
416975	1706276	2A02671	Esteem™+ 1-pc Drainable Pouch with Durahesive® Plus Adhesive Technology
416976	1706278	2A00585	Esteem™+ 1-pc Drainable Pouch with Durahesive® Plus Adhesive Technology
416976	1706278	2A00586	Esteem™+ 1-pc Drainable Pouch with Durahesive® Plus Adhesive Technology
416976	1706278	2A04528	Esteem™+ 1-pc Drainable Pouch with Durahesive® Plus Adhesive Technology
416976	1706278	2A04529	Esteem™+ 1-pc Drainable Pouch with Durahesive® Plus Adhesive Technology
416976	1706278	2A04530	Esteem™+ 1-pc Drainable Pouch with Durahesive® Plus Adhesive Technology
400598	1050402	2A03487	Active Life® One-Piece Drainable Pouch
400598	1050402	2B00300	Active Life® One-Piece Drainable Pouch
400599	1050403	2A03486	Active Life® One-Piece Drainable Pouch
400599	1050403	2B00301	Active Life® One-Piece Drainable Pouch

ConvaTec Inc is recalling the above item(s)/lot(s) due to a manufacturing error with the adhesive used in the manufacture specific batches of Esteem™+ 1-pc Drainable Pouch with Durahesive® Plus Adhesive Technology and Active Life® One-Piece Drainable Pouch. This issue may result in the user having difficulty in removing the release liner and therefore not be able to use the product. For this reason, all of the affected products should not be used.