

The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.

Product Name	Catalog Number	Batch No.	Exp. Date	Shelf Box Qty.
BD Ultra-Fine™ Pen Needles 32G x 4mm (5/32')	320122	9352052	20241231	100

BD is conducting a recall of the BD Ultra-Fine Pen Needles referenced in the table above. This recall is being conducted due to a small percentage of the product shelf cartons being incorrectly labeled as products intended for the Latin American market. Although the product description on the label “32G x 4mm pen needles” is correct, information pertinent to users in the US regarding compatibility with specific pen needles is missing. This defect is isolated to the specified Catalog and Lot Number listed in the table above. As the product itself is not defective there are no expected health consequences. BD distributed the affected lot between June 3, 2020 and July 20,2020.