

The Board of Pharmacy has received notice of the following product recall:

Becton, Dickinson and Company (BD) is recalling the following product:

Jamshidi™ Bone Marrow Biopsy/Aspiration Needle

Product Code	Lot Number
DJ4011X	0001303256
DJ4011X	0001303257

BD has identified that the product code / lot number combinations listed in the table above may be at risk of having some packaging pouches that may not be sealed properly.

The Board of Pharmacy strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.