

The Board of Pharmacy has received notice of the following product withdrawal:

**LEVEL OF NOTIFICATION:** Pharmacy/Hospital

**SUPPLIER:** Johnson & Johnson

Description	Lot #	NDC	UPC
HEALON GV PRO 0.85ML	UE31560, UE31521, UE31478, UE31440, UE31214, UE31164	5047465165	35047465165

Johnson & Johnson is voluntarily recalling the above lots due to complaints that the device may be difficult to remove from the eye, and potential clogging of phacoemulsification equipment leading to delay in the procedure or possible injury. This recall is to the pharmacy/hospital level.

The Board of Pharmacy strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.