

The Board of Pharmacy has received notice of the following medical device correction:

Description	Lot #	NDC	UPC
ACCU-CHEK AVIVA+ TST STRIP 50	498315	65702040710	36570240710
AVIVA+ T/S HLTH NWK-AC NFRS 50	497864	65702043810	36570243810

Roche is issuing an Urgent Medical Device Correction for the above items due to the inability of test strips to be dosed as intended. This recall is to the consumer level. Affected product started shipping November 2018.

The Board of Pharmacy strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.