

Pharmacy_Subscriberlist@DCA

From: Board of Pharmacy <pharmacy.subscriberlist@DCA.CA.GOV>
Sent: Friday, September 30, 2016 4:12 PM
To: PHARM-GENERAL@DCALISTS.CA.GOV
Cc: Pharmacy_Subscriberlist@DCA
Subject: Recall notice - Bausch & Lomb Inc.

Bausch & Lomb is recalling the item(s)/lot(s) below due to internal testing that showed simulated normal product use. Over the course of the recommended 35 uses of the lens case, testing showed occasional occurrences where the residual peroxide in the lens case after neutralization was above product specification. This recall is to the Healthcare Professional level.

Please immediately check your inventory, quarantine, discontinue distribution of the affected product, and contact your wholesaler for directions.

Description	Lot #	UPC
PEROXICLEAR 3% HYDRO PEROX 2x12OZ	All LOTS	31011903817
PEROXICLEAR 3% HYDRO PEROX 2x12OZ	ALL LOTS	31101903908
PEROXICLEAR 3% HYDRO PEROX 3OZ SAMPLE KIT	ALL LOTS	31011909712
PEROXICLEAR 3% HYDRO PEROX12OZ	ALL LOTS	31011903816
PEROXICLEAR 3% HYDRO PEROX12OZ	ALL LOTS	31011903910
PEROXICLEAR 3% HYDRO PEROX3OZ	ALL LOTS	31011903815