

To: DC Managers
From: Cynthia Gaw
Date: November 9, 2010
RC: 10-201

URGENT!!! DRUG RECALL!!! URGENT!!!

FDA/SUPPLIER CLASS OF RECALL: Not Yet Classified
LEVEL OF NOTIFICATION: McKesson Customer
SUPPLIER: AstraZeneca # 02837

Description	Lot #	Exp Date	NDC	UPC	Econo #
SYMBICORT MDI 160/4.5MCG 120DS	3000225C00	09/2011	00186037020	30186037020	1773100
	3000225D00	09/2011			
	3000228C00	09/2011			

AstraZeneca is voluntarily recalling the above lots because dials may stop advancing during actuations on some inhalers. This recall is to the McKesson Customer level. Product started shipping May 2010.

Please check your inventory for the above lots. Complete and return the Business Reply Card even if you do not have the affected lots. If you have the recalled lots, quarantine immediately. Return these lots with the Packing Slip using the prepaid UPS Return Service shipping label to the address below. Please contact Stericycle at 1-877-496-5042 if you need a Business Reply Card.

Stericycle
2670 EXECUTIVE DRIVE, STE A
INDIANAPOLIS, IN 46241

McKesson customers are to check their inventory and return to Stericycle.

Please contact Stericycle at 1-877-496-5042 if you have additional questions. Contact AstraZeneca if you have medical questions.

PLEASE NOTE: Customers currently participating in a McKesson administered Return to Vendor program should return this product to their designated returns processor. All others should follow the instructions provided by the manufacturer.

This recall is being conducted with the knowledge of the FDA.

Information contained in this document was provided by AstraZeneca.

FOR McKESSON USE ONLY:

- Please follow Recall Guidelines as outlined in the Reclamation SOP for a CLASS II RECALL.
- PROCESS TO THE RETAIL LEVEL.