

The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.

Description	Lot #	NDC
CHEMO+ GOWN L USX N/S CT5503	All lot numbers from 1901XXK to 2010XXK	90315055034
CHEMO+ GOWN XL USX N/S CT5504	All lot numbers from 1901XXK to 2010XXK	
CHEMO+ GOWN XL USX N/S DP5002G	All lot numbers from 1901XXK to 2010XXK	
CT5055 ChemoPlus Chemo Gown Open-Back Large	All lot numbers from 1901XXK to 2010XXK	
CT5502 ChemoPlus Chemo Gown - Medium	All lot numbers from 1901XXK to 2010XXK	
DP5001G ChemoPlus Chemo Gown Open Back X-Large	All lot numbers from 1901XXK to 2010XXK	
DP5004G ChemoPlus Chemo Gown - XXL Large	All lot numbers from 1901XXK to 2010XXK	

Cardinal is recalling the above items/lots due to potential rips, holes, tears, incomplete and/or open seams, staining, foreign material or debris on the gowns. This recall is to the health care professional level. Affected product started shipping January 1, 2019.