

The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.

Product Name	Catalog (Ref) No.	Lot No.	Expiration Date (YYYYMMDD)	UDI	Product Package Size
BD SafetyGlide™ 3ml 25G x 1	305924	8289599	2023-09-30	50382903059244	400
				30382903059240	50
				00382903059249	1

BD is conducting a medical device recall of BD SafetyGlide Syringes referenced in the table above. The recall is being conducted due to the syringes having scale markings printed in varying degrees, resulting in missing and/or partially printed scale numbers and scale lines. Although this defect would most likely be noticeable to the user, minimizing the impact to patient health, in a worst-case scenario, it could result in a potential mis-dose volume of up to 0.5ml. This defect is isolated to the specified Catalog and Lot Number listed in the table above. BD distributed the affected lot between November 14, 2018 and May 15, 2019.