

The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.

Description	Product Numbers	Part Numbers	Batch Numbers
No-Sting SKIN-Prep Wipes and No-Sting SKIN-Prep Swabs	59420600,	59420600	71430, 71660, 71860, 72090, 72210, 72330, 72780, 73020, 73030, 73320, 73420, 73600, 73850, 74020, 74240, 74640, 74890, 75220, 75340
	59420700,	59420700	65630, 65850, 65980, 66210, 66520, 66670, 72790
	66800712,	66800712	65350
	66800787,	66800787	67340
	66800788,	66800788	67660, 67750, 67770, 67860, 73460
	66800789,	66800789	73820
	66800790	66800790	66790

Smith & Nephew Medical Limited has initiated a field action to remove certain batches of No-Sting SKIN-Prep Wipes and No-Sting SKIN-Prep Swabs due to a manufacturing error resulting in the presence of acetic acid causing a vinegar-like odor and potential minor skin irritation. The recalled batches were distributed between June 2022 and May 2023.