

The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.

<b>Description</b>	<b>Lot #</b>	<b>Exp Date</b>	<b>NDC</b>	<b>UPC</b>
NEEDL SPC BD 30GX1/2 100	9128897	06/30/2024	08290305106	38290305106

BD is recalling the above item/lot due to presence of blood on the cannulas of the product noted, resulting in possible delay in treatment. This recall is to the retail level. Affected product started shipping August 27, 2019.