

The Board of Pharmacy has received notice of the following product market withdrawal:

**Description: BED RAIL HAND STND ESS P1410 Lot# and Exp Dates: ALL LOTS**

**UPC: 75475661410**

**Econo: 1735752**

**Description: BED RAIL HAND STND ESS P1410DS Lot# and Exp Dates: ALL LOTS**

**UPC: 75475661410**

**Econo: 2108413**

Essential Medical Supply is voluntarily recalling all lots of the above items due to users can become entrapped within the rails of the product or between the rail and mattress. This recall is to the Consumer level. Affected product started shipping October 1, 2006.

The Board of Pharmacy strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.