

The Board of Pharmacy has received notice of the following product recall:

Description: EXACTAMED O/DSP 5ML AMB8105 1C Lot #s: B304N045P, B301N070P, B301N017P

NDC: 17191008105

UPC: 31719108105

Econo# 1712843

Description: EXACTAMED O/DSP 5ML AMB8505 5C Lot #s: B303N108P, B302N077P, B301N049P, B301N023P

NDC: 17191008505

UPC: 31719108505

Econo# 1712926

Description: EXACTAMED ORAL 5mL, Baxter Code: H93884205 Lot #s: B307N015P, B303N042P, B301N056P

NDC: n/a

UPC: n/a

Econo# n/a

Baxter is issuing a safety alert for the above items/lots due to the potential for the 3mL dispenser to be misinterpreted or not recognized based on the package labeling and used as a 5 mL dispenser. This could result in medication errors.

The Board of Pharmacy strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.