

The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required. The purpose of this is to inform you that B. Braun Medical Inc. (BBMI) is issuing a pharmaceutical recall for or specific batches of S8506, STERILE WATER FOR INJECTION USP 3L; S8705, 70% DEXTROSE INJECTION USP, 2L; R8306, LACTATED RINGER'S IRRIGATION USP,3L, and R8206, 0.90% SODIUM CHLORIDE IRR USP, 3L, due to the potential for leakage from the diaphragm port once the foil is removed.

The depth of this recall is to the hospital/healthcare facility level. This recall is being conducted with the knowledge of the United States Food & Drug Administration.

<b>Product Catalog Number</b>	<b>NDC Number</b>	<b>Product Description</b>	<b>Expiration Date</b>	<b>Region Distributed</b>	<b>Lot Number</b>	<b>Distribution Range</b>
S8506	0264-7385-60	STERILE WATER FOR INJECTION USP, 3L	31 MAY 2028	US	J5H165	13Aug2025-02Dec2025
S8506	0264-7385-60	STERILE WATER FOR INJECTION USP, 3L	30 APR 2028	US	J5E237	11Jun2025-23Oct2025
S8705	0264-7387-50	70% DEXTROSE INJECTION USP, 2L	30 NOV 2026	US	J5H166	22Aug2025-10Nov2025
					J5H167	02Jul2025-14Nov2025
					J5E226	24Jul2025-31Jul2025
					J5E227	25Jun2025-22Sep2025
					J5E228	05Jun2025-03Sep2025
				US	J5E229	10Jul2025-

R8206	0264-7388-60	0.9% SODIUM CHLORIDE IRR USP, 3L	30 APR 2028		J5E230	21Aug2025 28Jul2025- 20Aug2025
					J5E231	28Jul2025- 11Sep2025
					J5E232	09Jun2025- 04Aug2025
					J5E233	03Jun2025- 10Nov2025
					J5E238	01Aug2025- 05Dec2025
					J5E239	05Jun2025- 11Jun2025
					J5E240	06Jun2025- 05Dec2025
					J5E241	05Aug2025- 26Oct2025
					J5E242	09Jul2025- 07Aug2025
R8206	0264-7388-60	0.9%SODIUM CHLORIDE IRR USP, 3L	30 APR 2028	US	J5E243	06Aug2025- 10Nov2025
					J5E244	06Aug2025- 08Aug2025

					J5E245	09Jun2025-14Nov2025
					J5E247	06Aug2025-11Aug2025
					J5E248	11Jun2025-11Aug2025
					J5H151	04Jun2025-08Aug2025
					J5H152	03Jun2025-18Sep2025
					J5H153	17Jun2025-08Aug2025
R8206	0264-7388-60	0.9% SODIUM CHLORIDE IRR USP, 3L	31 MAY 2028	US	J5H154	06Aug2025-13Aug2025
					J5H155	11Aug2025-01Dec2025
R8306		LACTATED RINGER'S IRRIGATION USP,3L		US	J5E234	12Jun2025-03Dec2025

	0264-7389-60		30 APR 2028		J5E236	24Jul2025-14Aug2025
					J5H158	04Aug2025-08Oct2025
					J5H159	08Aug2025-22Dec2025
					J5H160	20Aug2025-27Jan2026
					J5H161	20Aug2025-05Sep2025
R8306	0264-7389-60	LACTATED RINGER'S IRRIGATION USP,3L	31 MAY 2028	US	J5H162	05Sep2025-26Sep2025
					J5H163	26Jun2025-17Dec2025
					J5H174	19Sep2025-05Feb2026
					J5H175	07Aug2025-28Jan2025