

The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.

The purpose of this letter is to advise you that Cardinal Health has initiated a medical device product recall for the Monoject™ 1mL Luer Lock Syringe, Tuberculin Print, Sterile. The outer packaging and blister pack labeling identify the product as 1mL Luer Lock Tuberculin Syringe (Product Code 1180100777), however the syringes contained within the packages are U-100 Insulin syringes.

A U-100 Insulin syringe in a 1mL Luer Lock Tuberculin package may pose a potential risk of incorrect dose administration, especially in the pediatric/neonatal population due to these syringes being used to administer small dose drugs. However, the likelihood of this risk is low due to volume equivalency between the syringes, clinician familiarity with dose equivalency, and visible differences between the graduations on a 1mL Tuberculin syringe (left) and a U-100 Insulin syringe (right) as shown in the images of the barrel printing below. The U-100 Insulin syringes also state “USE U-100 INSULIN ONLY” (highlighted below) as an additional safeguard to prevent inadvertent usage with other drugs.

Cardinal Health has not received any reports of harm or adverse events.

Product Code	Product Description	Lot Numbers	UDI
1180100777	Monoject™ 1mL Luer Lock Syringe, Tuberculin Print, Sterile	531164X	10884521013414 (each)
		531162X	20884521013411 (box)
			30884521013418 (case)