The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.

| Description | Lot # and Exp Date | NDC | UPC |
|--------------------------------|--------------------|-------------|-------------|
| SYR SAL 12CC FILL 10CC KEN 180 | ALL LOTS | 08881570121 | 38881570121 |
| SYR SAL 12CC FILL 3CC KEN 180 | ALL LOTS | 08881570123 | 38881570123 |
| SYR SAL 12CC FILL 5CC KEN 180 | ALL LOTS | 08881570125 | 38881570125 |

Cardinal Health is recalling *all lots* of the above items due to the potential for the plunger to draw back after the air has been expelled and reintroduce air back into the syringe. This recall is to the retail level. Affected product started shipping July 1, 2019.