

The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.

Description	Lot & Exp Date	NDC	UPC
Accu-Chek Aviva Plus 10ct test strips		65702040510	Not Stocked by McKesson
ACCU-CHEK AVIVA+ TST STRIP 50		65702040710	36570240710
ACCU-CHEK AVIVA+ TST STRIP 100	ALL LOTS;	65702040810	36570240810
AVIVA+ T/S HLTH NWK-AC NFRS 50		65702043810	36570243810
Accu-Chek Aviva Plus 50ct Mailorder strips	Product has an 18 month shelf life.	65702043610	Not Stocked by McKesson
ACCU-CHEK SMRTVIEW TST STRP 50		65702049210	36570249210
ACCU-CHEK SMARTVW 100 TST STRP		65702049310	36570249310
Accu-Chek SmartView 50ct Mailorder strips		65702049510	Not Stocked by McKesson

Roche is issuing a Medical Device Correction on *all lots* of the above items due to complaints about test strip vials opening while still inside a sealed carton during shipment. Per the notice from Roche, this Medical Device Correction is to be communicated with customers who have received these products in the last 18 months. This correction is to the consumer level.