

The Board of Pharmacy has received notice of the following product recall:

Cardinal Health is voluntarily recalling the below list of items/lots due to a manufacturing defect that was found for the cartridge component, which could compromise the sterility barrier of the product. This recall is to the retail pharmacy level. Affected product started shipping January 28, 2019.

Description	Lot #	NDC	UPC
NEEDL SPEC KEN 14G 1200011 100	904218	08881200011	38881200011
	904918		
	906329		
NEEDL MNJC 15GX1.5 1200029 100	904956	08881200029	38881200029
NEEDL SPEC KEN 16G 1200037 100	904212	08881200037	38881200037
	905612		
	907002		
NEEDL SPEC KEN 16G 1200045 100	905613	08881200045	38881200045
NEEDL SPEC KEN 18G 1200078 100	907007	08881200078	38881200078
NEEDL BEV KEN 27G 1200508 100	904901	08881200508	38881200508
Monoject 20 G x 1-1/2" Bulk	905657	08881200664	38881200664
	906327		
	906341		
Monoject Vet Pak 22 ga x 1-1/2"	909117	08881201456	38881201456
Monoject Vet Pak 20 ga x 1-1/2"	905611	08881201498	38881201498
Monoject Vet Pak 19 ga x 1-1/2"	905662	08881201522	38881201522
Monoject Vet Pak 18 ga x 1-1/2"	904219	08881201548	38881201548
	904220		
	907710		
Monoject Vet Pak 16 ga x 1-1/2"	902108	08881201654	38881201654
NEEDL BLNT KEN 15G 1202314 25	905659	08881202314	38881202314
	905660		
	906343		
NEEDL BLNT KEN 16G 1202322 25	902110	08881202322	38881202322
	902111		

Description	Lot #	NDC	UPC
Monoject Blunt Cannula 19G x 1-1/2"	905617	08881202355	38881202355
	905618		
	905619		
	905620		
	907004		
	907005		
NEEDL BLNT KEN 20G 1202363 25	907734	08881202363	38881202363
Monoject Blunt Cannula 22G x 1-1/2"	903530	08881202389	38881202389

The Board of Pharmacy strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.