

The Board of Pharmacy has received notice of the following product recall:

Description	Lot #	NDC	UPC
OPTIRAY 320 100ML PISYRGRFID20	L162A; L146DS; L141C	00019132300	30019132300
SOD CHL INJ 0.9% 125ML RFID 20	L169A; L159A	00019118827	30019118827

Guerbet is voluntarily recalling the above items/lots due to a recent software update affecting the products ability to be used properly. This recall is to the retail pharmacy level.

The Board of Pharmacy strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.